

Hepatitis B

anti-HBe

ANTIBODY TO HEPATITIS B VIRUS E ANTIGEN ELISA

One-step Incubation, Competition Principle

INSTRUCTIONS FOR USE

INTENDED USE

This kit is an enzyme-linked immunosorbent assay for qualitative detection of antibodies to hepatitis B virus e antigen (anti-HBe) in human serum or plasma. It is intended for use in clinical laboratories for diagnosis and management of patients related to infection with hepatitis B virus.

SUMMARY

Hepatitis B virus (HBV) is an enveloped, double-stranded DNA virus belonging to the Hepadnaviridae family and is recognized as the major cause of blood transmitted hepatitis together with hepatitis C virus (HCV). Infection with HBV induces a spectrum of clinical manifestations ranging from mild, inapparent disease to fulminant hepatitis, severe chronic liver disease which in some cases can lead to cirrhosis and carcinoma of the liver. Classification of a hepatitis B infection requires the identification of several serological markers expressed during three phases (incubation, acute and convalescent) of the infection. Now several diagnostic test are used for screening, clinical diagnosis and management of the disease.

The hepatitis B e antigen (HBeAg), which is a product of the pre-C/C gene, has been found in the hepatocytes during proliferation of the hepatitis B virus. After proteolysis, the HBe protein is secreted into the serum in size varying from 16 kD to 20 kD. The e-antigen is associated with high degree of viral proliferation. HBeAg appears shortly after HBsAg and is detectable from few days to several weeks after infection. After treatment, in the recovery phase following acute hepatitis B, HBeAg - the first serological marker - becomes negative and is replaced by the corresponding antibody (anti-HBe). If HBeAg is considered a specific marker of infectivity, the presence of anti-HBe antibody in blood is recognized to be a clinical sign of recovery from the infection. Acute and persistent HBV infections can also occur without HBeAg being detectable. Demonstration of anti-HBe in these patients is an indication of the presence HBV mutants. These may be associated with high, low, or non-detectable quantities of virus. For this, anti-HBe test is meaningful only in association with the HBeAg test for monitoring the course of an HBV infection.

PRINCIPLE OF THE ASSAY

This kit is based on solid phase, one-step incubation, competitive principle ELISA method. Anti-HBe, if present in the sample competes with purified anti-HBe conjugated to horseradish peroxidase (HRP-Conjugate) for a fixed amount of purified HBeAg pre-coated in the wells. When no anti-HBe presents in the sample, the HRP-conjugated anti-HBe will be bound with the antigens inside the wells and any unbound HRP-Conjugate is removed during washing. Chromogen A and B solutions are added into the wells and during incubation, the colorless Chromogens are hydrolyzed by the bound HRP-Conjugate to a blue-colored product. The blue color turns

yellow after stopping the reaction with sulfuric acid. No or low color developing suggests the presence of antibodies to HBeAg in the sample.

Assay principle scheme Competition ELISA

$Ag(p)+Ab(s)+(Ab)ENZ \rightarrow [Ag(p) - Ab(s)] \rightarrow \text{No color} \quad (+)$
 $Ag(p) + (Ab)ENZ \rightarrow [Ag(p) - (Ab)ENZ] \rightarrow \text{Blue} \rightarrow \text{Yellow Color} \quad (-)$

Incubation	Immobilized Complex	Coloring	Results
60 min.		15 min.	

Ag(p)–pre-coated HBeAg;
Ab(s)– anti-HBe in sample;
(Ab)ENZ– HRP conjugated anti-HBe ;

COMPONENTS



96 Tests

- **MICROWELL PLATE** 1plate
Blank microwell strips fixed on white strip holder. The plate is sealed in aluminium pouch with desiccant.
8 × 12/12 × 8-well strips per plate. Each well contains recombinant HBeAg. The microwell strips can be broken to be used separately. Place unused wells or strips in the plastic sealable storage bag together with the desiccant and return to 2–8°C.
- **NEGATIVE CONTROL** 1vial
Yellowish liquid filled in a vial with green screw cap.
1ml per vial.
Protein-stabilized buffer tested non-reactive for anti-HBe.
Preservatives: 0.1% ProClin 300.
Ready to use as supplied.
Once open, stable for one month at 2-8°C.
- **POSITIVE CONTROL** 1vial
Red-colored liquid filled in a vial with red screw cap.
1ml per vial.
anti-HBe diluted in protein-stabilized buffer.
Preservatives: 0.1% ProClin 300.
Ready to use as supplied.
Once open, stable for one month at 2-8°C.
- **HRP-CONJUGATE REAGENT** 1vial
Red-colored liquid filled in a white vial with red screw cap.
6.5ml per vial.
Horseradish peroxidase-conjugated purified anti-HBe.
Ready to use as supplied.
Once open, stable for one month at 2-8°C.
- **STOCK WASH BUFFER** 1bottle
DILUTE BEFORE USE
Colorless liquid filled in a clear bottle with white screw cap.
30ml per bottle.
PH 7.4, 20 × PBS (Contains Tween-20 as a detergent).
The concentration must be diluted 1 to 20 with distilled/deionized water before use. Once diluted, stable for one week at room temperature, or for two weeks when stored at 2-8°C.
- **CHROMOGEN SOLUTION A** 1vial
Colorless liquid filled in a white vial with green screw cap.
7ml per vial.
Urea peroxide solution.
Ready to use as supplied.
Once open, stable for one month at 2-8°C.
- **CHROMOGEN SOLUTION B** 1vial
Colorless liquid filled in a black vial with black screw cap.
7ml per vial.
TMB solution (Tetramethyl benzidine dissolved in citric acid).
Ready to use as supplied.
Once open, stable for one month at 2-8°C.
- **STOP SOLUTION** 1vial

Colorless liquid filled in a white vial with white screw cap.
7ml per vial.

Diluted sulfuric acid solution (2.0M H₂SO₄).

Ready to use as supplied.

- **PLASTIC SEALABLE BAG** 1unit

For enclosing the strips not in use.

- **CARDBOARD PLATE COVER** 1sheet

To cover the plates during incubation and prevent evaporation or contamination of the wells.

- **PACKAGE INSERTS** 1copy

ADDITIONAL MATERIALS AND INSTRUMENTS REQUIRED BUT NOT PROVIDED

1. Freshly distilled or deionized water.
2. Disposable gloves and timer.
3. Appropriate waste containers for potentially contaminated materials.
4. Disposable V-shaped troughs.
5. Dispensing system and/or pipette (single or multichannel), disposable pipette tips.
6. Absorbent tissue or clean towel.
7. Dry incubator or water bath, 37±0.5°C.
8. Microshaker for dissolving and mixing conjugate with samples.
9. Microwell plate reader, single wavelength 450nm or dual wavelength 450nm and 630nm.
10. Microwell aspiration/wash system.

SPECIMEN COLLECTION, TRANSPORTATION AND STORAGE

1. **Sample Collection:** Either fresh serum or plasma samples can be used for this assay. Blood collected by venipuncture should be allowed to clot naturally and completely – the serum/plasma must be separated from the clot as early as possible as to avoid hemolysis of the RBC. Care should be taken to ensure that the serum samples are clear and not contaminated by microorganisms. Any visible particulate matters in the sample should be removed by centrifugation at 3000 RPM for at least 20 minutes at room temperature, or by filtration on 0.22µ filters. Plasma samples collected into EDTA, sodium citrate or heparin may be tested, but highly lipaemic, icteric, or hemolyzed samples should not be used as they could give erroneous results in the assay. Do not heat inactivate samples. This can cause sample deterioration.
2. **Transportation and Storage:** Store samples at 2-8°C. Samples not required for assaying within 3 days should be stored frozen (-20°C or lower). Multiple freeze-thaw cycles should be avoided. For shipment, samples should be packaged and labeled in accordance with the existing local and international regulations for transport of clinical samples and ethological agents.

SPECIAL INSTRUCTIONS FOR WASHING

1. A good washing procedure is essential to obtain correct and precise analytical data.
2. It is therefore recommended to use a good quality ELISA microplate washer, maintained at the best level of washing performances. In general, no less than 5 automatic washing cycles with dispensing of 350-400µl/well, are sufficient to avoid false positive reactions and high background (all wells turn yellow).
3. To avoid cross-contaminations of the plate with sample or

HRP-conjugate, after incubation do not discard the content of the wells, but allow the plate washer to aspirate it automatically.

4. Anyway, we recommend calibrating the washing system on the kit itself in order to match the declared analytical performances. Assure that the microplate washer's liquid dispensing channels are not blocked or contaminated, and sufficient volume of Wash buffer is dispensed each time into the wells.
5. In case of manual washing, we suggest to perform at least 5cycles, dispensing 350-400µl/well and aspirating the liquid for 5times. If poor results (high background) are observed, increase the washing cycles or soaking time per well.
6. In any case, the liquid aspirated out the strips should be treated with a sodium hypochlorite solution(final concentration of 2.5%) for 24 hours, before liquids are dispensed in an appropriate way.
7. The concentrated Washing solution should be diluted **1 to 20** before use. For one plate, mix 30 ml of the concentrate with 570ml of water for a final volume of 600ml diluted Wash Buffer. If less than a whole plate is used, prepare the proportional volume of solution.

STORAGE AND STABILITY

The components of the kit will remain stable through the expiration date indicated on the label and package when stored between 2-8 °C ; **do not freeze**. To assure maximum performance of this anti-HBe ELISA kit, during storage protect the reagents from contamination with microorganism or chemicals.

PRECAUTIONS AND SAFETY

This kit is intended **FOR IN VITRO USE ONLY** 

FOR PROFESSIONAL USE ONLY

The ELISA assay is a time and temperature sensitive method. To avoid incorrect result, strictly follow the test procedure steps and do not modify them.

1. Do not exchange reagents from different lots, or use reagents from other commercially available kits. The components of the kit are precisely matched as to achieve optimal performance during testing.
2. Make sure that all reagents are within the validity indicated on the kit box and are of the same lot. Never use reagents beyond the expiry date stated on reagents labels or on the kit box.
3. **CAUTION - CRITICAL STEP:** Allow the reagents and samples to stabilize at room temperature(18-30°C) before use. Shake reagent gently before, and return to 2-8°C immediately after use.
4. Use only sufficient volume of sample as indicated in the procedure steps. Failure to do so, may cause in low sensitivity of the assay.
5. Do not touch the bottom exterior of the wells; fingerprints or scratches may interfere with microwell reading.
6. When reading the results, ensure that the plate bottom is dry and there are no air-bubbles inside the wells.
7. Never allow the microplate wells to dry after the washing step. Immediately proceed to the next step. Avoid the formation of air bubbles when adding the reagents.
8. Avoid assay steps long time interruptions. Assure same working conditions for all wells.

9. Calibrate the pipette frequently to assure the accuracy of samples/reagents dispensing. Always use different disposal pipette tips for each specimen and reagents as to avoid cross-contaminations. Never pipette solutions by mouth.
10. The use of automatic pipettes is recommended.
11. Assure that the incubation temperature is 37°C inside the incubator.
12. When adding samples, avoid touching the well's bottom with the pipette tip.
13. When reading the results with a plate reader, it is recommended to determine the absorbance at 450nm or at 450nm with reference at 630nm.
14. All specimens from human origin should be considered as potentially infectious.
15. Materials from human origin may have been used in the kit. These materials have been tested with tests kits with accepted performance and found negative for antibodies to HIV ½, HCV, TP and HBsAg. However, there is no analytical method that can assure that infectious agents in the specimens or reagents are completely absent. Therefore, handle reagents and specimens with extreme caution as if capable of transmitting infectious diseases. Strict adherence to GLP (Good Laboratory Practice) regulations can ensure the personal safety. Never eat, drink, smoke, or apply cosmetics in the assay laboratory.
16. Bovine derived sera may have been used in this kit. Bovine serum albumin (BSA) and fetal calf sera (FCS) are derived from animals from BSE/TSE free-geographical areas.
17. The pipette tips, vials, strips and sample containers should be collected and autoclaved for 1hour at 121°C or treated with 10% sodium hypochlorite for 30minutes to decontaminate before any further steps for disposal.
18. The Stop solution (2M H₂SO₄) is a strong acid. Corrosive. Use it with appropriate care. Wipe up spills immediately or wash with water if come into contact with the skin or eyes. ProClin 300 used as a preservative can cause sensation of the skin.
19. The enzymatic activity of the HRP-conjugate might be affected from dust, reactive chemical, and substances like sodium hypochlorite, acids, alkalins etc. Do not perform the assay in the presence of such substances.
20. Materials Safety Data Sheet (MSDS) available upon request.
21. If using fully automated microplate processing system, during incubation, do not cover the plates with the plate cover. The tapping out of the remainders inside the plate after washing, can also be omitted.

ASSAY PROCEDURE

- Step 1 Reagents preparation:** Allow the reagents and samples to reach room temperature (18-30°C) for at least 15-30minutes. Check the Wash buffer concentrate for the presence of salt crystals. If crystals have formed in the solution, resolubilize by warming at 37°C until crystals dissolve. Dilute the stock Wash Buffer 1 to 20 with distilled or deionized water. Use only clean vessels to dilute the buffer.
- Step2 Numbering Wells:** Set the strips needed in strip-holder and number sufficient number of wells including three Negative controls (e.g. B1, C1, D1) two Positive controls (e.g. E1, F1) and one Blank (e.g. A1, neither samples nor HRP-Conjugate should be added into the Blank well). If the results will be

determined by using dual wavelength plate reader, the requirement for use of Blank well could be omitted. Use only number of strips required for the test.

- Step3 Adding Sample and HRP-Conjugate:** Add 50µl of Positive control, Negative control, and Specimen into their respective wells. **Note: Use a separate disposal pipette tip for each specimen, Negative and Positive Control as to avoid cross-contamination.** Add 50µl HRP-Conjugate to each well except into the Blank and mix by tapping the plate gently.
- Step4 Incubating:** Cover the plate with the plate cover and incubate for **60 minutes at 37°C**. It is recommended to use thermostat-controlled water tank to assure the temperature stability and humidity during the incubation. If dry incubator is used, do not open the door frequently.
- Step5 Washing:** At the end of the incubation, remove and discard the plate cover. Wash each well 5times with diluted Wash buffer. Each time allow the microwells to soak for 30-60seconds. After the final washing cycle, turn the plate onto blotting paper or clean towel, and tap it to remove any remaining liquids.
- Step6 Coloring:** Dispense 50µl of Chromogen A and 50µl Chromogen B solution into each well including the Blank. Incubate the plate at 37°C for 15minutes avoiding light. The enzymatic reaction between the Chromogen solutions and the Conjugate produces blue color in Negative control and anti-HBe negative sample wells.
- Step7 Stopping Reaction:** Using a multichannel pipette or manually, add 50µl Stop Solution into each well and mix gently. Intensive yellow color develops in Negative control and anti-HBe negative sample wells.
- Step8 Measuring the Absorbance:** Calibrate the plate reader with the Blank well and read the absorbance at 450nm. If a dual filter instrument is used, set the reference wavelength at 630nm. Calculate the Cut-off value and evaluate the results (**Note:** read the absorbance within 5minutes after stopping the reaction).

INTERPRETATION OF RESULTS AND QUALITY CONTROL

Each microplate should be considered separately when calculating and interpreting results of the assay, regardless of the number of plates concurrently processed. The results are calculated by relating each sample's optical density (OD) value to the Cut-off value (C.O.) of the plate. If the Cut-off reading is based on single filter plate reader, the results should be calculated by subtracting the Blank well OD value from the print report values of samples and controls. In case the reading is based on dual filter plate reader, do not subtract the Blank well OD from the print report values of samples and controls.

1. Calculation of Cut-off value

Cut-off value (C.O.) = *Nc × 0.5

*Nc = the mean absorbance value for three negative controls.

Example: of Cut-off calculation:

1. Calculation of Nc			
Well No:	B1	C1	D1
egative controls OD value	1.727	1.731	1.729
c Nc=	1.729		
2. Calculation of Cut-off : Cut-off (C.O.)= 1.729 × 0.5 =0.864			

If one of the Negative control values does not meet the Quality control range specifications, it should be discarded, and the mean value is calculated again using the remaining two values. If more than one Negative control OD value does not meet the Quality control range specifications, the test is invalid and must be repeated.

2. Quality control range:

The test results are valid if the Quality Control criteria are verified. It is advisable that each laboratory must establish appropriate quality control system with quality control material similar to or identical with the patient sample being analyzed.

1. The OD value of the Blank well, which contains only Chromogens and Stop solution, is less than 0.080 at 450 nm.
2. The OD value of the Negative control must be equal to or greater than 0.800 at 450/630nm or at 450nm after blanking.
3. The OD value of the Positive control must be less than 0.100 at 450/630nm or at 450nm after blanking.

3. Interpretations of the results:

(S = the individual absorbance (OD) of each specimen

Negative Results (S/C.O. >1): Samples giving an absorbance greater than the Cut-off value are considered negative, which indicates that no antibodies to HBV e antigen have been detected using this anti-HBe ELISA kit. This result should not be used alone to establish the infection state.

Positive Results (S/C.O. ≤1): Samples giving absorbance less than, or equal to the Cut-off value are initially reactive for this assay, which indicates that antibodies to HBV e antigen have probably been detected with this anti-HBe ELISA kit. Any reactive samples should be retested in duplicates. Repeatedly reactive samples could be considered positive for anti-HBe. This result should not be used alone to establish the infection .

Borderline: Samples with absorbance to Cut-off ratio between 0.9 and 1.1 are considered borderline samples and retesting is recommended. Repeatedly positive samples could be considered positive for anti-HBc.

TEST PERFORMANCE AND EXPECTED RESULTS

Analytical Specificity:

No cross reactivity observed with samples from patients infected with HAV, HCV, HIV, CMV, and TP. No interference from rheumatoid factors up to 2000U/ml observed during clinical testing. The assay performance characteristics are unaffected from elevated concentrations of bilirubin, hemoglobin, and triolein. Frozen specimens have been tested to check for interferences due to collection and storage.

Clinical Specificity: The clinical specificity of this kit has been evaluated by a panel of samples obtained from 980 healthy blood donors and 270 undiagnosed hospitalized patients. Any positive results were confirmed with another commercially available anti-HBe ELISA kit. The repeatedly reactive samples and samples confirmed positive with the reference test were not included in the calculation of specificity.

Specificity Samples	-	+	Confirmed	Specificity	False positives	
Blood donors	980	955	25	24	99.89%	1
Hospital	270	250	20	9	99.60%	1
TOTAL	1250	1205	45	33	99.74%	2

Clinical Sensitivity: To calculate the clinical sensitivity of the kit, a panel of sample obtained from 654 hepatitis B patients was used. Each patient clinical status was established based upon reference assays for detection of HBsAg, HBeAg, anti-HBs, anti-HBe, and anti-HBc.

Sensitivity Samples	-	+	Confirmed	Sensitivity	False Negatives	
Acute	367	71	296	96	100%	0
Chronic	72	3	69	69	100%	0
Recovery	215	51	164	164	100%	0
TOTAL	654	125	529	329	100%	0

Reproducibility		Within Run		Between Run	
Specimen Type	No tests	Mean OD	CV%	Mean OD	CV%
Weak positive	10	0.639	5.8%	0.645	6.4%
Moderate positive	10	0.427	7.4%	0.445	8.0%
Strong positive	10	0.011	21%	0.015	22%
Negative control	10	2.106	4.8%	2.128	4.9%

LIMITATIONS

1. Non-repeatable positive result may occur due to the general biological characteristics of the ELISA method. In very rare cases, some HBV mutants or subtypes could remain undetectable. Antibodies may be also undetectable during the early stages of the disease and in some immunosuppressed individuals.
2. If, after retesting of the initially reactive samples, the assay results are negative, these samples should be considered as non-repeatable (false positive) and interpreted as negative. As with many very sensitive ELISA assays, false positive results can occur due to the several reasons, most of which are related but not limited to inadequate washing step.
3. Any positive results must be interpreted in conjunction with the patient clinical information and other laboratory results.
4. Common sources for mistakes: kits beyond the expiry date, bad washing procedures, contaminated reagents, incorrect assay procedure steps, insufficient aspiration during washing, failure to add samples or reagents, equipment, timing, volumes, sample nature and quality.
5. The prevalence of the marker will affect the assay's predictive values.
6. This kit is intended ONLY for testing of individual serum or plasma samples. Do not use it for testing of cadaver samples, saliva, urine or other body fluids, or pooled (mixed) blood.
7. This is a qualitative assay and the results cannot be used to measure antibodies concentrations.

INDICATIONS OF INSTABILITY OR DETERIORATION OF THE REAGENTS

1. Values of the Positive or Negative controls, which are out of the indicated Quality control range, are indicator of possible deterioration of the reagents and/or operator or equipment errors. In such case, the results should be considered as invalid and the samples must be retested. In case of constant erroneous results classified as due to deterioration or instability of the reagents, immediately substitute the reagents with new ones.
2. If after mixing of the Chromogen A and B solutions into the wells, the color of the mixture turns blue within few minutes, do not continue carrying out the testing and

replace the reagents with fresh ones.

VALIDITY

Please do not use this kit beyond the expiry date indicated on the kit box and reagent labels.

REFERENCES:

1. Takahashi, K. et al. (1976). Association of Dane particles with e antigen in the serum of asymptomatic carriers of hepatitis B surface antigen. J. Immunol., 117, 102
2. Cappel, R. et al. (1977). E antigen and antibody DNA polymerase and inhibitors of DNA polymerase in acute and chronic hepatitis. J. Infect. Dis., 136, 617.
3. A new antigen-antibody system. Clinical significance in long-term carriers of hepatitis B surface antigen. J. Am. Med. Assoc., 231, 356. Bruss V, et al.

SUMMARY OF THE MAJOR COMPONENTS OF THE KIT:

Microwell plate	One/ 96 wells
Negative/Positive control	One each/ 1ml
HRP-Conjugate	One/ 6.5ml
Wash Buffer	One/30ml
Chromogen A/b, Stop solution	One each/ 7ml

Note: the components of individual kits are not interchangeable

SUMMARY OF THE ASSAY PROCEDURE:

Add Sample	50µl
Add HRP-Conjugate	50µl
Incubate	60minutes
Wash	5times
Coloring	50µl A + 50µl B
Incubate	15minutes
Stop the reaction	50µl stop solution
Read the absorbance	450nm or 450/630 nm

Example of controls/samples dispensing scheme

	1	2	3	4	5	6	7	12
A	Blank	S3								
B	Neg.	...								
C	Neg.	...								
D	Neg.									
E	Pos.									
F	Pos.									
G	S1									
H	S2									